

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100415

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/19/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.81: 0.079
LOT #: 02511080A1
EXPIRATION: 02/01/2013

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 17:40 |
| INTERNAL STANDARD | VERIFIED | 17:40 |
| EXTERNAL STANDARD | 0.079 | 17:40 |
| BLANK TEST | 0.000 | 17:41 |
| EXTERNAL STANDARD | 0.078 | 17:41 |
| BLANK TEST | 0.000 | 17:42 |
| EXTERNAL STANDARD | 0.078 | 17:42 |
| BLANK TEST | 0.000 | 17:43 |
| EXTERNAL STANDARD | 0.078 | 17:43 |
| BLANK TEST | 0.000 | 17:44 |
| EXTERNAL STANDARD | 0.078 | 17:44 |
| BLANK TEST | 0.000 | 17:45 |

Average = 0.0782
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

| | |
|-------------------------|--------|
| Sample Chamber = 49.1°C | PASSED |
| Breath Tube = 45.9°C | PASSED |

PUMP INFO

| | |
|-----------------------|--------|
| Flow Rate = 5.246 L/M | PASSED |
|-----------------------|--------|

DETECTOR INFO

| | |
|----------|--------|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |


FILTER INFO

| | |
|----------|--------|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

| | |
|-------------------|--------|
| INTERNAL STANDARD | PASSED |
|-------------------|--------|


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9 day of Nov, 20 11


Nikki Roth, Notary Public
My Commission Expires With Office



BMB
10/21/11